



Certificate of Calibration

Jan 1, 2023

This certificate applies to the following devices or series of devices as furnished by Silicon Labs.

Si7005	Si7006	Si7007
Si7013	Si7020	Si7021
Si7022	Si7023	Si7050
Si7051	Si7053	Si7054
Si7055		

In order to insure the datasheet specifications and any agreed-upon customer specific requirements are met, each sensor is calibrated under controlled temperature and relative humidity conditions.

This calibration is completed using equipment calibrated on a regular basis by a laboratory certified to operate in accordance with ISO/IEC 17025, is a member of the International Laboratory Accreditation Cooperation (ILAC) and is a signatory of the ILAC Mutual Recognition Arrangement (MRA).

The temperature and humidity standards (transfer standards) used for calibration are traceable via these accredited laboratories to standards set by the National Institute of Standards and Technology (NIST).

Each device has undergone functional testing following calibration.

A handwritten signature in blue ink that reads 'Kathy Torres'.

Kathy Torres

Product Quality Manager

400 West Cesar Chavez, Austin TX 78701

Phone: (512) 416-8500